



INTERNATIONAL STANDARD ISO 10723:1995
TECHNICAL CORRIGENDUM 1

Published 1998-07-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Natural gas — Performance evaluation for on-line analytical systems

TECHNICAL CORRIGENDUM 1

Gaz naturel — Évaluation des performances des systèmes d'analyse en ligne

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 10723:1995 was prepared by Technical Committee ISO/TC 193, *Natural gas*, Subcommittee SC 1, *Analysis of natural gas*.

Page 6

Subclause 6.2

In the equation, replace s by y_s so that the equation reads:

$$x_s = \frac{y_s}{y_{std}} \times x_{std}$$

Correct the definition of x_{std} and y_{std} to read:

x_{std} is the concentration in the standard;

y_{std} is the response to the standard.

It should be noted that the expression “standard” as used here, and elsewhere in the document (e.g. subclause 6.3), means the calibration gas.

ICS 75.060

Ref. No. ISO 10723:1995/Cor.1:1998(E)

Descriptors: natural gas, gas analysis, test equipment, specifications, performance.